

Manual Tensile Specimen Gauge Length Meter



(Picture just for reference)

Usage and Characteristics

It is a specialized equipment designed according to the regulations of GB/T228. It can meet the customers' different needs of gauge, which is the necessary auxiliary instrument of laboratory.

Advantages

Easy operation and high precision

Main Technical Parameters

Model	KASON-DX5
Control methods	Manual
Max. gauge length	300mm
Point space	5mm or 10mm (adjustable)
Striking point each time	Single point 1
Accuracy	+/-0.1mm
Outline size	About 500*300*250mm
Net Weight	About 22Kg



Authentication Certificate of Quality Management System

Registration No.19922Q01072R0S

This is to certify that

Jinan Kason Testing Equipment Co., Ltd.

Unified Social Credit Code: 9137010458994939X0

Registered/Production Address: No. 4715, Jingshi West Road,
Huaiyin District, Jinan City, Shandong Province

Office Address: Room 2-1108, Harmony Square, Huaiyin District, Jinan City

Is in conformity with :

GB/T 19001-2016/ISO 9001:2015

The certificate is valid for the following scope :

Assembly and sales of testing machines

This certificate should be used effectively with the legal requirements of the administrative license, qualification license, compulsory certification, which the certification scope involved.

In the case that the organization regularly receives surveillance assessments, the certificate shall be valid when used together with the Quick Response Code.

Date of Issue: 2022-07-19

Date of Expiry: 2025-07-18

Date of Initial Certification: 2022-07-19

General Manager



中国认可
国际互认
管理体系
MANAGEMENT SYSTEM
CNAS C182-M

China Standard Huaxin (Beijing)
Certification Center Co.,Ltd.



Address of Certification Authority: 5 Floor, YuQuan Building, No.3 Shijingshan Road, Shijingshan District, Beijing Tel: 010-88255986 P.C.: 100049

Note: The detailed information is available on the Certification and Accreditation Administration of the People's Republic of China official website (www.cnca.gov.cn) the authority's website(www.cshcc.cn) and the Quick Response Code on the left.



MACHINERY DIRECTIVE ATTESTATION OF CONFORMITY

Technical file of the company mentioned below has been inspected and audit has been completed successfully.

2006/42/EC Machinery Directive Annex VIII has been taken as references for these processes.

Company Name : **Jinan Kason Testing Equipment Co., Ltd.**
 Company Address : No.4715, Jingshi West Road, Huaiyin District, Jinan, Shandong, China
 Related Directives and Annex : **2006/42/EC Machinery Directive/Annex VIII**
 Related Standards : **EN ISO 12100:2010; EN 60204-1:2018**
 Product Name : **Testing Machine**
 Report No and Date : TCF-JKTE230605-MD
 Product Brand/Model/Type :

WDS-1, WDS-2, WDS-3, WDS-5, WDS-10, WDS-20, WDS-30, WDS-50, WDS-100, WDS-200, WDS-300, WDW-1E, WDW-2E, WDW-3E, WDW-5E, WDW-10E, WDW-20E, WDW-30E, WDW-50E, WDW-100E, WDW-200E, WDW-300E, WDW-500E, WDW-600E, WDW-1000E, ETM103, ETM203, ETM303, ETM503, ETM104, ETM204, ETM304, ETM504, ETM105, WE-300D, WE-600D, WE-1000D, WE-2000D, HUT305S, HUT605S, HUT105S, HUT126S, WEW-300D, WEW-600D, WEW-1000D, WEW-1200D, WEW-2000D, WEW-3000D, HUT305E, HUT605E, HUT105E, HUT126E, HUT-A306E, HUT-A206E, WAW-300D, WAW-600D, WAW-1000D, WAW-1200E, WAW-2000E, WAW-3000, WAW-2000A, WAW-3000A, WAW-4000A, HUT-A206W, HUT-A306W, HUT-A406W, HUT305W, HUT605W, HUT105W, HUT126W, WAL-LAW, HUT-L, JB-300, JB-300B, JB-500B, JBS-300, JBS-500, JBW-300, JBW-500, JBDW-300, JBDW-500, CST-50, VU-2Y, VU-1S, DWC-40, DWC-60, DWC-80, DWC-100, DWC-196, YES-1000, YES-2000, YES-3000, YAW-1000, YAW-300, YAW-2000, YAW-3000, YAW-5000, NDS-200, NDS-500, NDS-1000, NDS-2000, NDW-200, NDW-500, NDW-1000, NDW-2000, NDW-3000, NDW-5000, NDW-10000, EZ-1, EZ-3, EZ-10, JWJ-10, TLS, TLW, HWS, PLS-20, PLS-50, PLS-100, PLS-300, PLS-500, PLS-1000, KASON-1, RDJ-10, RDJ-20, RDJ-50, RDJ-100, DWTT-20000, DWTT-30000, DWTT-40000, DWTT-50000, DWTT-80000, DWTT-100000, PTF-302, PTF-302M, PTF-502, PTF-502M, PTF-302T, PTF-502T, PTF-302W, PTF-302TW, PTF-502W, PTF-502TW, PTF302CW, PTF452CW, PTF602CW, PTF752CW, PTF302CT, PTF452CT, PTF602CT, PTF752CT, PTF302W, PTF452W, PTF602W, PTF752W, PTF302CW, PTF452CW, PTF452CT, PTF752CW, PTF752CT, PTF302C, HCT106, HCT206, HCT306, KSCIM, HYJ, HEZ-3, HEZ-5, HEZ-10, HEZ-20, KS8T-10, KASON 370.02, KASON 370.05, KASON 370.10, KASON 370.25, KASON 370.50, MMW, GBW-60B, GBS-60B, HUT, ETM, XNR, GW-40, GW-50, MRS, MMU, MRH, GBW, GW, LSW, HBT, GLY, GCJ-1000, GWQ-1000, GNZ-1000, GQR-1000, GJR-300, GSS-1000, GMS-100, GTW-500, GQC-5, GGG-100, HR-150A, HRD-150, HRS-150, XHR, HB, HVS, HV, HBRV, W-20, HBA-1, LHT, Q-ZA, MP-2, XQ-2B, XD-30M, 4XCE, 4XB, JSZ, MOPAO, MPT, P, MP, BG, LBG-32, MPJ-35, Q-Z, Q-100B, SQ, Q-3A, QG-4, LSQ, LDQ, GTQ-5000A, GTQ-5000B, DTQ, ZXQ, 4XC, HST, KASON, KS, KSCUT, HTR, HTPR, HTB, HSTR-45S, HTV, HTMV, HTBRV, METRPRESS, MPJ

Certificate Number : **M.2023.206.C86313**
 Initial Assessment Date : 07.06.2023
 Registration Date : 08.06.2023
 Reissue Date/No : -
 Expiry Date : **07.06.2028**


 UDEM International Certification
 Auditing Training Centre Industry
 and Trade Inc. Co.

The validity of the certificate can be checked through www.udem.com.tr. Upon completion of EC declaration of conformity it is used solely at the manufacturer's responsibility. This certificate remains the property of UDEM International Certification Auditing Training Centre Industry and Trade Inc. Co. to whom it must be returned upon request. The above named firm must keep a copy of this certificate for 15 years from the registration of certificate. This certificate only covers the product(s) stated above and UDEM must be noticed in case of any changes on the product(s)

Address: Muthukent Mahallesi 2073 Sokak (Eski 93 Sokak) No:10 Çankaya - Ankara - TURKEY

Phone: +90 0312 443 03 90 Fax: +90 0312 443 03 76

E-mail: info@udem.com.tr www.udem.com.tr



FOCUS IN MATERIAL TEST

KASONTTEST®

JINAN KASON TESTING
EQUIPMENT Co., LTD.

Duandian Industrial Park , Jingshi Road, Jinan City, China.

P: +86 159 10081986

E: admin@jnkason.com | W: www.syjlab.com

